



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

## LIMITED TENDER ENQUIRY

No: IAV/303/PUR/2022

Sealed quotations are invited for the supply of the following stores:

SI No	Name of the Item	Quantity
1	Magnetic stirrer with hot plate (Specification as per Schedule of Supply)	2

- The envelopes containing the quotation should bear the superscription “Purchase of Magnetic stirrer with hot plate” and should be addressed to “The Director, Institute of Advanced Virology, Bio 360 Life Sciences Park, Thonnakkal, Thiruvananthapuram, Kerala – 695 317”
- The tenderers may submit the quotations on their own papers/letter head.
- Last date for receipt of quotations is 15/03/2022, 12:00pm
- Late quotations will not be accepted.
- The quotations will be opened on 15/03/2022 at 01:00pm in the presence of such of the tenderers or their authorized representatives who may wish to be present at that time.
- The maximum period required for delivery of the articles should also be mentioned.
- Details of the requirements and the conditions governing their supply can be obtained free on request from the office till 14/03/2022.

Director  
IAV, Thiruvananthapuram

Place: Thiruvananthapuram

Date: 09/03/2022



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

## TERMS AND CONDITIONS

Quotation Number	<b>IAV/303/PUR/2022</b>
Due date and time for receipt of quotations	<b>15/03/2022, 12:00pm</b>
Date and time for opening of quotations	<b>15/03/2022, 01:00pm</b>
Date up to which the rates are to remain firm for acceptance	<b>15/06/2022</b>
Designation and address of officer to whom the quotation is to be addressed	The Director Institute of Advanced Virology Bio 360 Life Sciences Park Thonnakkal Thiruvananthapuram, Kerala – 695 317
Superscription: Quotation for purchase of Magnetic stirrer	

Sealed quotations are invited for the supply of the materials specified in the schedule attached below/overleaf. The rates quoted should be for delivery of the articles at the place mentioned below the schedule. The necessary superscription, due date for the receipt of quotations, the date up to which the rates will have to remain firm for acceptance and the name and address of officer to whom the quotation is to be sent are noted above. Any quotation received after the time fixed on the due date is liable to be rejected. The maximum period required for delivery of the articles should also be mentioned.

Quotations not stipulating period of firmness and with price variation clause and/or 'subject to prior sale' condition are liable to be rejected.



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

The acceptance of the quotations will be subject to the following conditions:

1. Acceptance of the quotation constitutes a concluded contract. Nevertheless, the successful tenderer must within a fortnight/a month after the acceptance of his quotation furnish 5 per cent of the amount of the contract as security deposit and execute an agreement at his own cost for the satisfactory fulfilment of the contract, if so required.
2. Withdrawal from the quotation after it is accepted or failure to supply within a specified time or according to specifications will entail cancellation of the order and purchases being made at the offerers expenses from elsewhere, any loss incurred thereby being payable by the defaulting party. In such an event the Institute reserves also the right to remove the defaulter's name from the list of Institute suppliers permanently or for a specified number of years.
3. Samples, duly listed, should be forwarded if called for under separate cover and the unapproved samples got back as early as possible by the offerers at their own expenses and the Institute will in no case be liable for any expenses on account of the value of the samples or their transport charges, etc. In case, the samples are sent by railway; the railway receipt should be sent separately, and not along with the quotation since the quotation will be opened only on the appointed day and demurrage will have to be paid if the railway parcels are not cleared in time. Quotations for the supply of materials are liable to be rejected unless samples, if called for of the materials tendered for are forwarded. The approved samples may or may not be returned at the discretion of the undersigned. Samples sent by V.P.Post or "freight to pay" will not be accepted.
4. No representation for enhancement of price once accepted will be considered during the currency of the contract.
5. Any attempt on the part of tenderers or their agents to influence the Officers concerned in their favour by personal canvassing will disqualify the tenderers.
6. If any license or permit is required, tenderers must specify in their quotation and also state the authority to whom application is to be made.
7. The quotation may be for the entire or part supplies. But the tenderers should be prepared to carry out such portion of the supplies included in their quotation as may be allotted to them.
8. (a) In cases where a successful tenderer, after having made partial supplies fails to fulfil the contracts in full, all or any of the materials not supplied may, at the discretion of the competent



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

authority be purchased by means of another tender/quotation or by negotiation or from the next higher tenderer who had offered to supply already and the loss, if any, caused to the Institute shall thereby together with such sums as may be fixed by the Government towards damages be recovered from the defaulting tenderer.

(b) Even in cases where no alternate purchases are arranged for the materials not supplied, the proportionate portion of the security deposit based on the cost of the materials not supplied at the rate shown in the tender of the defaulter shall be forfeited and balance alone shall be refunded.

(c) Any sum of money due and payable to the contractor (including Security Deposit returnable to him) under this contract may be appropriated by the authorities of the Institute or any other person authorized by the Institute and set-off against any claim of the authorities of the Institute for the payment of a sum of money arising out of or under any other contract made by the contractor with the Purchasing Officer or Institute or any other person authorized by the Institute.

9. The prices quoted should be inclusive of all taxes, duties, delivery charges, installation charges cesses, etc., which are or may become payable by the contractor under existing or future laws or rules of the country of origin/supply or delivery during the course of execution of the contract. The item has to be delivered and installed within 30 days from date of supply order. The LD clause will be applicable as per Institute rules for late supply

10. (a) Ordinarily payments will be made only after the supplies are actually verified and taken to stock but in exceptional cases, payments against satisfactory shipping documents including certificates of Insurance will be made up to 90 per cent of the value of the materials at the discretion of Institute. Bank charges incurred in connection with payment against documents through bank will be to the account of the contractor. The firms will produce stamped pre-receipted invoices in all cases where payments (advance/final) for release of railway receipts/shipping documents are made through Banks. In exceptional cases where the stamped receipts of the firms are not received for the payments (in advance) the unstamped receipt of the Bank (i.e. counterfoils of pay-in-slips issued by the Bank) alone may be accepted as a valid proof for the payment made.



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

(b) The tenderers shall quote also the percentage of rebate (discount) offered by them in case the payment is made promptly within fifteen days/within one month of taking delivery of stores.

11. Any sum of money due and payable to the successful tenderer or contractor from Institute shall be adjusted against any sum of money due to the Institute from him under any other contracts.

12. Special conditions, if any, printed on the quotation sheets of the tenderer or attached with the tender will not be applicable to the contract unless they are expressly accepted in writing by the purchases.

Director  
IAV, Thiruvananthapuram

Place: Thiruvananthapuram

Date: 09/03/2022



# INSTITUTE OF ADVANCED VIROLOGY

An Autonomous Institution under Science and Technology Department, Government of Kerala

## SCHEDULE OF SUPPLY

Sl No	Name of the Item	Quantity
1	Magnetic stirrer with hot plate	2

## SPECIFICATIONS OF MAGNETIC STIRRER WITH HOT PLATE

1. Magnetic stirrer should have speed up to 1,500rpm.
2. Magnetic stirrer should have temperature range of 500°C or more.
3. Magnetic stirrer should have surface area heating of 7.25 x 7.25 in.
4. Magnetic stirrer should have surface area heating of 18.4 x 18.4cm.
5. Top Plate Material should be Ceramic.
6. Magnetic stirrer should be cCSAus, CE and RoHS Compliant.
7. Magnetic stirrer should include 1 power cord.
8. Magnetic stirrer should have shipping weight of 5kg or more.
9. Magnetic stirrer should have stirring range of 50 to 1,500rpm.
10. Magnetic stirrer should have dimensions of (L x W x H): 13 x 8.2 x 3.8 in. (33 x 20.8 x 9.7cm).
11. Electrical Requirements should be 230V 50/60Hz.
12. Plug should be Indian type.
13. Magnetic stirrer should have maximum temperature (Metric) of 500°C.
14. Magnetic stirrer should have warranty of 3 years.
15. Stir-Trac features innovative engineering for slow-speed stirring, consistent speed control, and strong magnetic coupling.
16. Magnetic stirrer should have temperature adjustable in 1-degree increments.
17. Magnetic stirrer should have a Hot Top warning system, which protects from accidental burns with prominent display when heating surface is above 50°C.
18. Magnetic stirrer should have LED display for heating and stirring.
19. Magnetic stirrer should have an automatic over-temperature protection.
20. Magnetic stirrer should include Stir Bar.